



*Republic of the Philippines*  
**ZAMBOANGA CITY WATER DISTRICT**  
Pilar Street, Zamboanga City

**Addendum No. 2018-037**

This Addendum is issued to modify or amend the Bid Documents in view of the Pre-Bid Conference last **July 26, 2018**. This shall form an integral part of the Bid Documents. The Bids & Awards Committee hereby issues the following amendments, clarifications & updates:

PR No.	Item Description	Issues/Amendments/Updates/Clarifications for the particular Purchase Requisition	Issues/Amendments/Updates /Clarifications to the Bidding Documents
18-0210	Supply & Delivery of Biosafety Cabinets (Level 2) for laboratory use	<ul style="list-style-type: none"><li>There shall be no preference as to Country of Origin as provided in ITB Clause 7 of PBD</li><li>Filter: Ultra Low Particulate Air (ULPA)</li><li>Below are the revised technical specifications for Biosafety Cabinets:<ul style="list-style-type: none"><li>Hood Width (in): 48-50</li><li>Exterior Width (in): 52-56</li><li>Exterior Height (in): 61-63</li><li>Exterior Depth (in): 30-34</li><li>From Interior Depth (in): 49-51 to Interior Width (in): 46-51</li><li>From Interior Width (in): 29-31 to Interior Height (in): 25-31</li><li>Interior Depth (in): 18-24</li><li>Power (VAC): 220-240</li><li>Power (Hz): 50 or 60</li><li>Exhaust Volume (CFM): at least 339</li></ul></li><li>Delivery Period: Ninety (90) Calendar Days upon receipt of P.O.</li><li>Winning bidder shall provide the following Accessories:<ul style="list-style-type: none"><li>Computer Chair</li><li>Suitable AVR</li></ul></li><li>Please see attached revised Technical Specifications and Schedule of Requirements</li></ul>	<ul style="list-style-type: none"><li>Please refer to <b>GPPB Resolution No. 26-2017</b> regarding Deferment of the implementation of the mandatory submission of Philippine Government Electronic Procurement System (PhilGEPS) Certificate of Registration and Membership in Competitive Bidding</li><li>As per Section 25.2. of the 2016 Revised IRR of RA 9184, the first envelope shall contain the following technical information/documents, at the least:<ul style="list-style-type: none"><li>a) For the procurement of Goods:<ul style="list-style-type: none"><li>i) PhilGEPS Certificate of Registration and membership in accordance with Section 8.5.2 of this IRR, except for foreign bidders participating in the procurement by a Philippine Foreign Service Office or Post, which shall submit their eligibility documents under Section 23.1 of this IRR: Provided , That the winning bidder shall register with the PhilGEPS in accordance with Section 37.1.4 of this IRR.</li><li>ii) Statement of all Ongoing Government and Private Contracts;</li><li>iii) Statement of SLCC;</li><li>iv) NFCC Computation or committed Line of Credit;</li><li>v) JVA or the Duly Notarized Statement in accordance with Section 23.1(b) of this IRR, if applicable;</li><li>vi) Bid security in the prescribed form, amount and validity period;</li><li>vii) Technical Specifications, which may include production/delivery schedule, manpower requirements, and/or after-sales service/parts, if applicable;</li><li>viii) Omnibus Sworn Statement in accordance with Section 25.3 of this IRR; and</li><li>ix) For foreign bidders claiming eligibility by reason of their country's extension of reciprocal rights to Filipinos, a certification from the relevant government office of their country stating that Filipinos are allowed to participate in their government procurement activities for the same item or product.</li></ul></li></ul></li></ul>
18-0144	Supply & Delivery of Flash Drive, External Hard Drive, Memory Card, Internal Hard Drive, USB Keyboard & Continuous Ink Printer for various ZCWD Department/Divisions	<ul style="list-style-type: none"><li>Flash Drive, 4GB Capacity shall be revised to 8GB Capacity</li></ul>	
18-0149	Supply & Delivery of 1,800 bags of Portland Cement for use in fabrication of concrete meter stand for year 2018 to be carried in stock	<ul style="list-style-type: none"><li>There shall be no preference as to Country of Origin as provided in ITB Clause 7 of PBD</li><li>Payment will be for every delivery.</li></ul>	
18-0126	Supply & Delivery of C.I. Manhole Cover and Frame to be used for various line meter chambers		
18-0236	Supply & Delivery of Glassware Washer for laboratory use		<ul style="list-style-type: none"><li>As per Section 25.4. of the 2016 Revised IRR of RA 9184, The second envelope shall contain the financial information/documents as specified in the PBDs.</li><li>BAC shall adhere to the 2016 Revised IRR of RA 9184 and any discrepancies in the Instruction to Bidders, the 2016 Revised IRR of RA 9184 will prevail.</li><li>Clarification(s) on any part of the Bidding Documents or for an interpretation must be in writing and submitted to the Procuring Entity at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.</li></ul>

PR No.	Item Description	Issues/Amendments/Updates/Clarifications for the particular Purchase Requisition	Issues/Amendments/Updates /Clarifications to the Bidding Documents
			<ul style="list-style-type: none"><li>➤ All forms of bid security are acceptable however if the Bidder opts to submit a Surety Bond it must be callable on demand and must be accompanied by (1) Certificate of Authority from the Insurance Commission authorizing the Insurance/Finance Company to issue such bond and (2) another Certification for the surety bond from the Insurance Commission for bidding purpose of the specific project. A new form of bid security which is Bid Securing Declaration is acceptable, pursuant to GPPB Resolution No. 03-2012 dated January 27, 2012.</li><li>➤ Proper tabbing of bid documents must be observed by the bidders.</li><li>➤ The bidders are advised to observe the Bid Prices found in the ITB Clause 15.</li><li>➤ The bidders are advised to observe the Format and Signing of Bids found in the ITB Clause 19.</li><li>➤ The bidders are advised to observe the Sealing and Marking of Bids found in the ITB Clause 20.</li></ul>

For guidance and information of all concerned.

  
**MICHAEL ANGELO M. CARBON**  
Chairperson  
Bids and Awards Committee

Posting date: July 31, 2018  
Posted in the Phil-GEPS, ZCWD Website & Bid Bulletin

# Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Quantity	Total	Delivered, Weeks/Months
1	Biosafety Cabinets (Level 2)	1 unit		
	Filter: Ultra Low Particulate Air (ULPA)			
	Below are the revised technical specifications for Biosafety Cabinets:			
	Hood Width (in): 48-50			
	Exterior Width (in): 52-56			
	Exterior Height (in): 61-63			
	Exterior Depth (in): 30-34			
	From Interior Depth (in): 49-51 to Interior Width (in): 46-51			
	From Interior Width (in): 29-31 to Interior Height (in): 25-31			
	Interior Depth (in): 18-24			
	Power (VAC): 220-240			
	Power (Hz): 50 or 60			
	Exhaust Volume (CFM): at least 339			
	Winning bidder shall provide the following Accessories: - Computer chair - Suitable AVR			
	Delivery Period: Ninety (90) calendar days upon receipt of P.O.			
	Point of Delivery: ZCWD Water Treatment Plant, Pasonanca, Z.C.			



# Technical Specifications

Item	Specification		Statement of Compliance
			Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of <b>ITB</b> Clause 3.1(a)(ii) and/or <b>GCC</b> Clause 2.1(a)(ii).
Item No.	Description	Quantity	
1	Biosafety Cabinets (Level 2)	1 unit	
	Filter: Ultra Low Particulate Air (ULPA)		
	Below are the revised technical specifications for Biosafety Cabinets:		
	Hood Width (in): 48-50		
	Exterior Width (in): 52-56		
	Exterior Height (in): 61-63		
	Exterior Depth (in): 30-34		
	From Interior Depth (in): 49-51 to Interior Width (in): 46-51		
	From Interior Width (in): 29-31 to Interior Height (in): 25-31		
	Interior Depth (in): 18-24		

	Power (VAC): 220-240		
	Power (Hz): 50 or 60		
	Exhaust Volume (CFM): at least 339		
	Winning bidder shall provide the following Accessories: - Computer chair - Suitable AVR		