

# Technical Specifications

Item	Specification		Statement of Compliance
			Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of <b>ITB</b> Clause 3.1(a)(ii) and/or <b>GCC</b> Clause 2.1(a)(ii).
Item No.	Description	Quantity	
1	Rapid Test Kit for Total Coliform/E.Coli (100 Spls)	100 samples	
	Complete set of accessories except incubator		
	For Total and E. Coli Test for drinking Water based on the detection of B-D galactosidase and B-glucoronidase enzyme using chromogenic and flurogenic substrate in just 18 hours		
	Must be DOH approved Product		
	Expiration Period of Chemicals in Rapid Test Kit: One (1) Year from receipt of items		

# Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Quantity	Total	Delivered, Weeks/Months
1	Rapid Test Kit for Total Coliform/E.Coli (100 Spls)	100 samples		
	Complete set of accessories except incubator			
	For Total and E. Coli Test for drinking Water based on the detection of B-D galactosidase and B-glucoronidase enzyme using chromogenic and flurogenic substrate in just 18 hours			
	Must be DOH approved Product			
	Expiration Period of Chemicals in Rapid Test Kit: One (1) Year from receipt of items			
	Period of Delivery: 45 calendar days after receipt of P.O.			
	Point of Delivery: ZCWD Water Treatment Plant, Pasonanca, ZC			